



Overnight Pulse Oximetry Guide

Instructions for using Konica Minolta Pulseox-300i

- 1) Wear overnight oximeter with appropriate support (oxygen, NIV) as indicated
- 2) When ready to go to sleep put oximeter onto wrist (like a watch)



- 3) If needed probe can be secured with tape (e.g. tape the lead to back of the hand).
DO NOT TAPE OVER THE PROBE
- 4) Turn on by pressing button down, holding briefly, ensure screen is turned on



- 5) Wear all night until fully awake and ready to get up
- 6) If you notice it has come off during the night replace the probe
- 7) Do not turn oximeter off during the night, if you go to the toilet leave switched on and put it back on when you on return to bed